

## RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

DATE OF MANUFACTURE:	2017-12
KIT DESCRIPTION:	ALDOLASE (ALS)
CATALOGUE NO:	AD189
BATCH NO:	434403
EXPIRY DATE:	2019-11-28

SAMPLE	TARGET	RANGE	RESULT	UNITS
LEVEL II 408AD	9.27	6.95 – 11.60	8.70	U/L
LEVEL III 390AD	21.6	16.2 – 27.0	20.9	U/L

**APPROVED BY:** 

**QUALITY CONTROL REPRESENTATIVE** 

Server Danuelly