

CERTIFICATE OF ANALYSIS

Reagent: Microalbumin Multi-Calibrators
Catalogue Number: SE-252
Lot Number: 51374
Shelf Life: 12 months from date of manufacture
Expiry Date: 2018-06-30
Storage Conditions: 2-8°C

A. PHYSICAL CHARACTERISTICS

PARAMETER	SPECIFICATION
Performance	Values obtained for normal and abnormal level control sera are within the ranges established for the method.

B. CHEMICAL CHARACTERISTICS

Reagent Component

Microalbumin Multi-Calibrators

Major Active Ingredients

human albumin standards containing known quantities of human albumin

sodium chloride 150 mmol/L

Intended Use

For IN VITRO diagnostic use as a calibrator for the Sekisui microalbumin assay for quantitation of albumin in human urine.

The above reagent has passed all of our quality control procedures and has been released for distribution.

07-Aug-2017
Date

Z.M.D. Jao
Quality Control

CAT-COA-0147-2.0-Sekisui-MicroSE252