

RANDOX

CERTIFICATE OF ANALYSIS

This is to certify that the following kit was tested by a
Randox Testing Laboratory and approved by the QC Department.

DATE OF MANUFACTURE: April 2017
KIT DESCRIPTION: Immunoassay Premium Level 1
CATALOGUE NO: IA2638
LOT NUMBERS: 1666EC
EXPIRY DATE: 2019-10-28
SYSTEM OF ANALYSIS: Siemens Immulite 2000 XPI

TEST CONTROLS OR VIALS

ANALYTE	LOT NUMBER	RANGE	TARGET	RESULT	PRECISION (if applicable)
AFP (kU/l)	1666EC	8.4 - 12.6	10.5	11.0	N/A
				10.9	
TSH (mIU/l)	1666EC	0.112 -0.168	0.14	0.15	
				0.15	
TT3 (ng/dl)	1666EC	31.8 - 52.8	42.3	46.3	
				41.9	

QC TESTED BY: Davina Graham
DATE: 27 April 2017
RESULT (PASS/FAIL): PASS
QC APPROVED BY: Adele Martin.

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FIAE Man. R J Fitzgerald, W Fitzgerald OBE, R P Kelly BSc FCA.
Randox Laboratories Limited is a company registered in Northern Ireland, Company Number NI15738, VAT Number GB 353 0304 00.

FORM NO. 6521-QC
REVISION (4)
14 April 2016