



## RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

|                      |            |
|----------------------|------------|
| DATE OF MANUFACTURE: | 2018-07    |
| KIT DESCRIPTION:     | CK NAC     |
| CATALOGUE NO:        | CK3878     |
| BATCH NO:            | 457251     |
| EXPIRY DATE:         | 2019-07-28 |

| SAMPLE             | TARGET | RANGE     | RESULT | UNITS |
|--------------------|--------|-----------|--------|-------|
| LEVEL II<br>1237UN | 223    | 183 – 263 | 215    | U/L   |
| LEVEL III<br>766UE | 583    | 478 - 688 | 548    | U/L   |

APPROVED BY:

QUALITY CONTROL REPRESENTATIVE

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom  
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E [marketing@randox.com](mailto:marketing@randox.com) I [www.randox.com](http://www.randox.com)

Directors: Dr S F Fitzgerald BSc PhD DSc (Hons) FRAE MEd R J Fitzgerald W Fitzgerald OBE R P Kelly BSc FCA  
Randox Laboratories Limited is a company registered in Northern Ireland, Company Number NI15738, VAT Number GB 353 0304 00.