

RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

DATE OF MANUFACTURE:	2018-02
KIT DESCRIPTION:	Enzymatic Creatinine
CATALOGUE NO:	CR2336
BATCH NO:	438427
EXPIRY DATE:	2021-01-28

SAMPLE	TARGET	RANGE	RESULT	UNITS
Level II/ 892UN	133	106 – 160	137	μmol/L
Level III/ 766UE	379	304 – 454	394	μmol/L

APPROVED BY:

QUALITY CONTROL REPRESENTATIVE

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com