



## RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

|                      |            |
|----------------------|------------|
| DATE OF MANUFACTURE: | 2018-09    |
| KIT DESCRIPTION:     | CK-MB      |
| CATALOGUE NO:        | CK1296     |
| BATCH NO:            | 460428     |
| EXPIRY DATE:         | 2021-07-28 |

| SAMPLE                      | TARGET | RANGE   | RESULT | UNITS  |
|-----------------------------|--------|---------|--------|--------|
| CKMB<br>CONTROL /<br>4196CK | 162    | 138-186 | 149    | mmol/L |
| CKMB<br>CONTROL /<br>4077CK | 156    | 133-179 | 148    | mmol/L |

APPROVED BY:

QUALITY CONTROL REPRESENTATIVE

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom  
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E [marketing@randox.com](mailto:marketing@randox.com) I [www.randox.com](http://www.randox.com)

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FRAE MRB R J Fitzgerald W Fitzgerald OBE R P Kelly BSc FCA  
Randox Laboratories Limited is a company registered in Northern Ireland, Company Number NI15738, VAT Number GB 353 0304 00.

4706-QC Rev4  
27<sup>th</sup> September 2013