

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI) – Italia – Via delle Rose 10
Tel +39-0577-587111 Fax +39-0577-316690



Pag 1/1
Sede Legale e Amministrazione
20144 Milano (Italy) Via A. Solari 19 sc. 6
Tel. +39-02-4859121 r.a. Fax +39-02-48008530

QUALITY CONTROL CERTIFICATE

KIT NAME: SS-A - CHORUS

Code	86018		
Lot No.	363	2018-11	2017-08

REAGENTS

Label	Description	Lot.	Exp	No
D3181	SS-A	363	2018-11	6X6
B3181	CALIBRATORE SS-A	363	2018-11	1
C3181	CONTROL SERUM SS-A	363	2018-11	1

CONTROLS

Found results	Expected results
Calibrator (mOD)	649 ≥ 100
Control Serum (AU/mL)	29,0 22,5-47,5

CALIBRATOR

1	9	3	3	6	3	7	2	2	2	3	4	4	5	5	1	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature

19336372234456100
SS-A

CALIBRATOR 0.176 mL IV0

LOT 363

2018-11 2°C - 8°C

DIESSE Monteriggioni - Italy

CONTROL SERUM

1	9	3	3	6	3	7	1	6	2	9	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature

193363716296000001
SS-A

CONTROL+ 0.425 mL IV0

LOT 363

2018-11 2°C - 8°C

DIESSE Monteriggioni - Italy

The test has been carried out in accordance with the package insert

TESTED BY: Antonella Leo

APPROVED

RESPONSIBLE HEAD, Quality Control Reagents

DATE 03/11/2017

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested by an FDA approved method and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.