



RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

DATE OF MANUFACTURE:	2018-01
KIT DESCRIPTION:	CALIBRATOR LEVEL 3
CATALOGUE NO:	CAL2351
BATCH NO:	961UE
EXPIRY DATE:	2020-01-28

SAMPLE	ANALYTE	TARGET	RANGE	RESULT	UNITS
LEVEL III 961UE	ALBUMIN	29.2	N/A	29.5	G/L
LEVEL III 961UE	TOTAL PROTEIN	46.5	N/A	45.7	G/L

APPROVED BY:

QUALITY CONTROL REPRESENTATIVE

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FRAE MSc R J Fitzgerald W Fitzgerald OBE R P Kely BSc FCA
Randox Laboratories Limited is a company registered in Northern Ireland, Company Number NI15738, VAT Number GB 353 0304 00.