## **RAND@X**

## RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

| DATE OF MANUFACTURE: | 2017-01        |
|----------------------|----------------|
| KIT DESCRIPTION:     | FULL RANGE CRP |
| CATALOGUE NO:        | CP8315         |
| BATCH NO:            | 407287         |
| EXPIRY DATE:         | 2018-06-28     |

| SAMPLE               | TARGET | RANGE     | RESULT | UNITS |
|----------------------|--------|-----------|--------|-------|
| LEVEL I /<br>453LPC  | 23.8   | 19.0-28.6 | 22.7   | mg/L  |
| LEVEL II /<br>454LPC | 47.4   | 37.9-56.9 | 43.8   | mg/L  |

**APPROVED BY:** 

Lisa Ren.

QUALITY CONTROL REPRESENTATIVE