DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI) – Italia – Via delle Rose 10
Tel +39-0577-587111 Fax +39-0577-318690

Sede Legale e Amministrazione
20144 Milano (Italy) Via A. Solari 19 sc. 6
Tel. +39-02-4859121 r.a. Fax +39-02-48008530

QUALITY CONTROL CERTIFICATE

KIT NAME: RUBELLA IgM - CHORUS

Code 81031

Lot No. 296 **Expiry** 2017-12

REAGENTS

Label	Description	Lot.	Exp	No
D2241	RUBELLA IgM	296	2017-12	6X6
B2241	CALIBRATORE RUBELLA IgM	296	2017-12	1
C2241	CONTROL SERUM	296	2017-12	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	540	≥ 100
Control Serum (INDEX)	1,7	1,4-2,8
Negative Control (INDEX)	0,0	< 0,7



CALIBRATOR

1	3	5	2	9	6	6	0	0	0	0	0	4	5	0	0	0	0

Label checked with barcode reader **YES** ☒ Operator Signature .....

CONTROL SERUM

1	3	5	2	9	6	6	1	4	2	8	0	0	0	0	0	0	1

Label checked with barcode reader **YES** ☒ Operator Signature .....

The test has been carried out in accordance with the package insert

TESTED BY: Sara Benocci

APPROVED RESPONSIBLE HEAD,Quality Control Reagents.....

DATE 16/12/2016

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!
This kit may contain some reagents made with human serum or plasma.All serum or plasma used has been tested by an FDA approved method and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.