
TITLE	CERTIFICATE OF ANALYSIS – CELL-FREE DNA BCT CE
-------	--

CERTIFICATE OF ANALYSIS

The following released product satisfies Streck product specifications for product performance and appearance.

PRODUCT NAME: Cell-Free DNA BCT® CE

LOT #: 81620317

DATE OF MANUFACTURE: 2018-06-18

EXPIRATION DATE: 2020-07-02

ACTIVITY*(Specification Range 2.30 – 3.46): 2.86

QUALITY CONTROL APPROVAL: Barb Spencer
Print

QUALITY CONTROL APPROVAL: Barb Spencer
(Signature)

TITLE: QC Product Specialist

DATE: 2018-08-15

This product was manufactured under FDA QSR, ISO 9001/ISO 13485 and CMDR conditions.

This product is FOR EXPORT ONLY, not to be sold in the United States.

This product has been sterilized. The Sterilization Process has been established and validated in accordance with ISO 11137, Sterilization of Healthcare Products. All established dosage requirements are traceable to this nationally recognized standard.

*Reflects the active ingredient concentration in Cell-Free DNA BCT CE as measured via Streck's proprietary test method.