RANDOX

RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

| DATE OF MANUFACTURE: | 2018-12 |
|----------------------|------------|
| KIT DESCRIPTION: | GAMMA GT |
| CATALOGUE NO: | GT3874 |
| BATCH NO: | 471257 |
| EXPIRY DATE: | 2019-11-28 |

| SAMPLE | TARGET | RANGE | RESULT | UNITS |
|--------------------|--------|-----------|--------|-------|
| LEVEL II 1237UN | 61 | 52 – 70 | 62 | U/L |
| LEVEL III 766UE | 170 | 145 - 195 | 173 | U/L |

APPROVED BY:

QUALITY CONTROL REPRESENTATIVE