



RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

DATE OF MANUFACTURE:	2017-04
KIT DESCRIPTION:	Microalbumin Calibrator Series
CATALOGUE NO:	MA1567
BATCH NO:	415851
EXPIRY DATE:	2018-07-28

SAMPLE	TARGET	RANGE	RESULT	UNITS
Level I / 891MA	31.4	25.1-37.7	32.2	mg/l
Level II / 892MA	154.7	123.8-185.6	159.0	mg/l

APPROVED BY:

QUALITY CONTROL REPRESENTATIVE

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FRCR MRCP R J Fitzgerald W Fitzgerald OBE R P Kelly BSc FCA
Randox Laboratories Limited is a company registered in Northern Ireland. Company Number NI15738, VAT Number GB 353 0304 00.

4706-QC Rev4
27th September 2013