

Certificate of Analysis

Product Name: Reference Emulsion Fluid - Orig:USA
Product Number: 100240
Lot Number: 667457

Expiration Date: 10/31/2016
Storage: Room Temp

Specification	Specification Limit	Result
Coagulation Optics Test (ACL Index)	-8 - 8 UNITS	-8
Coagulation Control Level 1 PT	-0.9 - 0.9 S	-0.2
Coagulation Control Level 3 PT	-6.0 - 6.0 S	0.0
Coagulation Control Level 1 APTT	-3.0 - 3.0 S	-0.7
Coagulation Control Level 3 APTT	-9.0 - 9.0 S	-1.9
Normal Fibrinogen	-50 - 50 mg/dL	-6
Low Fibrinogen	-60 - 60 mg/dL	4

For in-vitro diagnostic use.

Storage Restriction: Do Not Freeze.

All results displayed above are expressed in comparison to the Master.



Quality Representative

Date: 2015-05-12