

DIESE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI) – Italia – Via delle Rose 10
Tel +39-0577-587111 Fax +39-0577-318690



Pag 1/1
Sede Legale e Amministrazione
20144 Milano (Italy) Via A. Solari 19 sc. 6
Tel. +39-02-4859121 r.a. Fax +39-02-48006530

QUALITY CONTROL CERTIFICATE

KIT NAME: SS-A - CHORUS

Code 86018		
Lot No. 361	2018-11	2017-08

REAGENTS

Label	Description	Lot.	Exp	No
D3181	SS-A	361	2018-11	6X6
B3181	CALIBRATORE SS-A	361	2018-11	1
C3181	CONTROL SERUM SS-A	361	2018-11	1

CONTROLS

Found results	Expected results
Calibrator (mOD)	546
Control Serum (AU/mL)	32,1
	≥ 100
	22,5-47,5

CALIBRATOR

1	9	3	3	6	1	7	2	2	2	3	4	4	5	5	1	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES

Operator Signature *Jo*

19336172234456100
SS-A

CALIBRATOR 0.175 mL IVD

LOT 361

2018-11 2°C - 8°C

DIESE Monteriggioni - Italy

CONTROL SERUM

1	9	3	3	6	1	7	1	6	2	9	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES

Operator Signature *JB*

19336171629000001
SS-A

CONTROL+ 0.425 mL IVD

LOT 361

2018-11 2°C - 8°C

DIESE Monteriggioni - Italy

The test has been carried out in accordance with the package insert

TESTED BY: Sara Benocci

APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... *Alvise*

DATE 30/08/2017

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested by an FDA approved method and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.