

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI) – Italia – Via delle Rose 10
Tel +39-0577-587111 Fax +39-0577-318690



Sede Legale e Amministrazione
20144 Milano (Italy) Via A. Solari 19 sc. 6
Tel. +39-02-4859121 r.a. Fax +39-02-48008530

CERTIFICATO DI CONTROLLO QUALITÀ/QUALITY CONTROL CERTIFICATE

PRODOTTO/PRODUCT ESR CONTROL CUBE

REF: 10435 (4x9 ml)
10436 (2x9 ml)

REAGENTI/REAGENTS

	Kit	Level 1	Level 2
Lot:	323	323	323
	2017-06	2017-06	2017-06
	2018-06	2018-06	2018-06

RISULTATI/RESULTS

	Level 1		Level 2		Temp.
	Risultato/ Result	Range	Risultato/ Result	Range	
VES-Matic Cube 30	7	2-12	84	59-109	OFF
VES-Matic Cube 80/VES-Matic Cube 200	6	1-12	60	35-85	OFF

Data/Date

13/07/2017

Testato da/Tested by:

Sara Benocci

Approvato da/Approved by:

Alessandra Cianchi (Responsabile CQ/QC Head)