



RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

DATE OF MANUFACTURE:	2016-02
KIT DESCRIPTION:	LIQUID UREA KINETIC METHOD
CATALOGUE NO:	UR3873
BATCH NO:	374249
EXPIRY DATE:	2017-04

SAMPLE	TARGET	RANGE	RESULT	UNITS
LEVEL II / 892UN	7.04	5.98-8.10	7.64	mmol/L
LEVEL III / 766UE	19.20	16.30-22.10	18.54	mmol/L

APPROVED BY:

QUALITY CONTROL REPRESENTATIVE

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FRC Path, R J Fitzgerald, W Fitzgerald OBE, R P Kelly BSc FCA
Randox Laboratories Limited is a company registered in Northern Ireland, Company Number NI15738, VAT Number GB 353 0204 00

4706-QC Rev4
27th September 2013