



RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

2018-01-28	2018-10
KIT DESCRIPTION:	FULL RANGE CRP
CATALOGUE NO:	CP3847
BATCH NO:	467471
EXPIRY DATE:	2021-01-28

SAMPLE	TARGET	RANGE	RESULT	UNITS
LEVEL II / 496LPC	44.2	35.4-53.0	40.1	mg/L
LEVEL III / 542LPC	68.9	55.1-82.7	64.1	mg/L

APPROVED BY:

QUALITY CONTROL REPRESENTATIVE

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FIAE Mar. R J Fitzgerald, W Fitzgerald CBE, R P Kelly BSc FCA.
Randox Laboratories Limited is a company registered in Northern Ireland. Company Number NI15738. VAT Number GB 353 0304 00.

4706-QC Rev4
27th September 2013