

ACON BIOTECH (HANGZHOU) CO.,LTD

CERTIFICATE OF ANALYSIS

PRODUCT NAME: Urinalysis Reagent Strips

PURCHASE NO.: 43040003

BATCH NO.: URS8010181

EXPIRATION DATE: 2020-02

QUANTITY: 200x100T

ANALYSIS		SPECIFICATIONS (Color card)	TEST RESULT
Glucose	Neg. urine	Negative	Complies
	250mg/dL	>100mg/dL <500mg/dL	Complies
Bilirubin	Neg. urine	Negative	Complies
	Yihua Urine controls	>++	Complies
Ketone	Neg. urine	Negative	Complies
	Yihua Urine controls	>40mg/dL <160mg/dL	Complies
S.G	1.000	$\geq 1.000 < 1.005$	Complies
	1.025	$> 1.020 < 1.030$	Complies
Blood	Neg. urine	Negative	Complies
	Yihua Urine controls	++~+++	Complies
pH	5.0	$\geq 5.0 < 6.0$	Complies
	9.0	$> 8.0 \leq 9.0$	Complies
Protein	Neg. urine	Negative	Complies
	Yihua Urine controls	>+, $\leq$ +++	Complies
Urobilinogen	Neg. urine	Negative	Complies
	4~12mg/dL	>4mg/dl ≤ 12 mg/dl	Complies
Nitrite	Neg. urine	Negative	Complies
	Yihua Urine controls	+	Complies
Leukocyte	Neg. urine	Negative	Complies
	Yihua Urine controls	>+ <+++	Complies

QC Supervisor: Sihong Zhang

Signature: Sihong Zhang

Date: Mar. 02, 2018