



CERTIFICATE OF ANALYSIS

This is to certify that the following kit was tested by a
Randox Testing Laboratory and approved by the QC Department.

DATE OF MANUFACTURE: October 2017
KIT DESCRIPTION: IA Premium Plus Tri-Level
CATALOGUE NO: IA3112
LOT NUMBERS: 432365 (1660EC, 1662EC, 1719EC)
EXPIRY DATE: 2019-10-28
SYSTEM OF ANALYSIS: Siemens Immulite 2000

TEST CONTROLS OR VIALS

ANALYTE	LOT NUMBER	RANGE	TARGET	RESULT	PRECISION (if applicable)
Prolactin (uU/ml)	1660EC	118-176	147	160	N/A
				159	
FSH (U/L)	1660EC	4.50-6.76	5.63	5.6	N/A
				5.7	
CEA (ug/l)	1660EC	4.42-6.62	5.52	5.6	N/A
				5.4	
Prolactin (uU/ml)	1662EC	518-778	648	666	N/A
				668	
FSH (U/L)	1662EC	19.9-29.9	24.9	23.6	N/A
				26.1	
CEA (ug/l)	1662EC	30.9-46.3	38.6	40.7	N/A
				40.4	
Prolactin (uU/ml)	1719EC	1138-1708	1423	1410	N/A
				1450	
FSH (U/L)	1719EC	49.0-73.4	61.2	65.3	N/A
				61.7	
CEA (ug/l)	1719EC	65.9-98.9	82.4	90.8	N/A
				84.8	

QC TESTED BY: Ilona Palmer
DATE: 12 Oct 2017
RESULT (PASS/FAIL): Pass

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FIAE Man. R J Fitzgerald, W Fitzgerald OBE, R P Kelly BSc FCA.
Randox Laboratories Limited is a company registered in Northern Ireland, Company Number NI15738, VAT Number GB 353 0304 00.

FORM NO. 6521-QC
REVISION (4)
14 April 2016



QC APPROVED BY:

Adele Martin.

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom
T +44 (0) 28 9442 2413 **F** +44 (0) 28 9445 2912 **E** marketing@randox.com **I** www.randox.com

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FIAE Man. R J Fitzgerald, W Fitzgerald OBE, R P Kelly BSc FCA.
Randox Laboratories Limited is a company registered in Northern Ireland, Company Number NI15738, VAT Number GB 353 0304 00.

FORM NO. 6521-QC
REVISION (4)
14 April 2016