

SeraQuest Certificate of Analysis

Product:

Anti-SM IgG

Catalog No.:

01-230

Lot No.:

A14018

Exp. Date:

2019-01-31

Reagent Blank Abs.

0.088

Calibrator 1 Abs.

1.962

Calibrator 2 Abs.

0.806

Microwell Plate:

Lot No. A14118

Exp. Date Jan/19

Accuracy Passed (QC Panel)

4.5

1.5

Conjugate:

Lot No. A14218

Exp. Date Jan/19

Negative Control:

Lot No. A14318

Exp. Date Jan/19

Index 0.2

Calibrator 1:

Lot No. A14418

Exp. Date Jan/19

Index

Calibrator 2:

Lot No. A14618

Exp. Date Jan/19

Index

Positive Control:

Lot No. A14518

Exp. Date Jan/19

Index 1.7 - 5.2

Human source material used in preparation of this product have been tested by FDA licensed methods and found to be non-reactive for HBsAg, Anti-HCV, Anti-HIV 1 and Anti-HIV 2. No known test method can offer assurance that Hepatitis B Virus, Hepatitis C Virus, Human Immunodeficiency Virus or other infectious agents are absent. All human blood-based products should be handled in accordance with good laboratory practices and current acceptable guidelines for the prevention of blood-based viral infections.

Prepared By: Hamilym

Date: /

Certified By: -260-yul 5

Date: 02/22

Reviewed By:

Date: 02/22/2018

QUEST INTERNATIONAL, INC.

8127 N. W. 29th Street Doral, Florida 33122

. .

Tel: 305-592-6991 Fax: 305-592-6834

Form QC 26 Rev. 1 Page 1 of 1